

Table 2

DENTAL CEPHALOMETRIC PATIENT EXPOSURE GUIDE

Maximum Entrance Skin Exposures (ESE) in mR

Projection	Part Thickness	Grid	SID	ESE
Lat. Skull	15 cm	No	66"	7-23

Guidance regarding Table 2:

- (1) The information in the above table is based upon data from CRCPD Publication E-03-2 "Patient Exposure and Dose Guide- 2003"
- (2) The median exposure is 15 mR, the 3rd quartile exposure is 23 mR and the 1st quartile exposure was computed to be 7 mR
- (3) 1999 NEXT data typical kVp range for this examination was 62 to 90 while the median kVp was 75.