

TABLE 5

FILTRATION REQUIRED versus OPERATING VOLTAGE

Total Filtration: Inherent + Added (in mm Al Equivalent)		
Operating Voltage (kVp)	Dental X-ray Systems	Other X-ray Systems
Below 50		0.5
50-70	1.5	1.5
> 70	2.5	2.5